

Thiobarbituric Acid Reactants (TBARS) Colorimetric Assay Kit

Catalog No: E-BC-K298-M

Method: Colorimetric method

Specification: 96T (Can detect 80 samples without duplication)

Instrument: Microplate reader

Sensitivity: 0.85 $\mu\text{mol/L}$

Detection range: 2.6-100 $\mu\text{mol/L}$

- ▶ Please kindly provide us the lot number (on the outside of the box) of the kit for more efficient service.

General information

▲ Intended use


This kit can be used to measure TBARS concentration in serum (plasma), animal tissue samples.

▲ Background

The organism produces oxygen free radicals through enzyme system and non-enzyme system, attacks polyunsaturated fatty acids in biofilm, induces lipid peroxidation, and thus forms lipid peroxide. Malondialdehyde (MDA) is one of the common products of lipid peroxidation in organisms. In clinical science, MDA is a biomarker of lipid peroxidation, which can reflect the degree of lipid peroxidation in organism and indirectly reflect the degree of cell injury.

▲ Detection principle

TBARS and TBA can react under high temperature and acid conditions and then form a pink compound, the concentration of which is linearly related to the concentration of TBARS in the sample. The TBARS concentration can be calculated by measuring the OD values at 530-540 nm.



▲ Kit components & storage

Item	Component	Specification	Storage
Reagent 1	Clarificant	12 mL × 1 vial	2-8°C , 6 months
Reagent 2	Acid Reagent	12 mL × 1 vial	2-8°C , 6 months
Reagent 3	TBA Reagent	Powder × 1 vial	2-8°C , 6 months, shading light
Reagent 4	200 μmol/L Standard	5 mL × 1 vial	2-8°C , 6 months
	Microplate	96 wells	No requirement
	Plate Sealer	2 pieces	

Note: The reagents must be stored strictly according to the preservation conditions in the above table. The reagents in different kits cannot be mixed with each other.

If you have any problem, please contact our Technical Service Center for help.
Phone: 240-252-7368(USA) Fax: 240-252-7376(USA)

Email: techsupport@elabscience.com

Please kindly provide us the lot number (on the outside of the box) of the kit for more efficient service.

▲ Materials prepared by users

Instruments

Microplate reader (530-540 nm), Vortex mixer, Micropipettor, Water bath

Consumptive material

Tips (10 μL, 200 μL, 1000 μL), EP tubes (1.5 mL, 2 mL, 10 mL)

Reagents

Double distilled water, Acetic acid, Normal saline (0.9% NaCl), PBS (0.01 M, pH 7.4)

▲ Safety data

Some of the reagents in the kit contain dangerous substances. It should be avoided to touch the skin and clothing. Wash immediately with plenty of water if touching it carelessly. All the samples and waste material should be treated according to the relevant rules of laboratory's biosafety.

▲ Precautions

Before the experiment, please read the instructions carefully, and wear gloves and work clothes.

▲ The key points of the assay

1. The temperature of water-bath and the time of incubation should be stabilized (95-100°C , 60 min)
2. In the incubation of 100°C water bath, the EP tube should not be closed directly. It is recommended to fasten the tube mouth with fresh-keeping film and make a small hole in the film.
3. The supernatant for colorimetric measurement should not contain sediment, otherwise it will affect the OD values. It is recommended to use a pipette to take the supernatant.

Pre-assay preparation

▲ Reagent preparation

1. Reagent 1 will solidify when stored at 2-8°C . Incubate the Reagent 1 at 37°C until transparent, liquid can be used.
2. Preparation of Reagent 2 application solution:
Mix 1.2 mL of Reagent 2 and 34 mL of double-distilled water fully. Prepare the fresh solution before use and it can be stored at 2-8°C for a day.
3. Preparation of Reagent 3 application solution:
Dissolve a vial of reagent 3 powder with 60 mL of double distilled water (90-100°C) and mix fully. Then add 60 mL of glacial acetic acid (self-prepared), mix fully and cool to room temperature. The prepared reagent 3 application solution can be stored at 2~8°C in the dark for 1 month.
4. Preparation of chromogenic agent:
Prepare the chromogenic agent according to the ratio of reagent 2 application solution: reagent 3 application solution =3: 1 (mix fully). Prepare the fresh solution before use and it must be use out in 24 hours.

▲ Sample preparation

The samples should be prepared as conventional methods. Also please refer to appendix II.

▲ Dilution of sample

It is recommended to take 2~3 samples with expected large difference to do pre-experiment before formal experiment and dilute the sample according to the result of the formal experiment and the detection range (2.6-100 $\mu\text{mol/L}$).

The recommended dilution factor for different samples is as follows (for reference only)

Sample type	Dilution factor
Human serum	1
Rat serum	1
10% Mouse brain tissue homogenization	1
Human serum	1
Rat serum	1
10% Rat liver tissue homogenization	1

Note: The diluent is normal saline (0.9% NaCl) or PBS (0.01 M, pH 7.4).

Assay protocol	
Ambient temperature	25-30°C
Optimum detection wavelength	532 nm

Instructions for the use of transferpette:

- (1) Equilibrate the pipette tip in that reagent before pipetting each reagent.
- (2) Don't add the liquid outside the tips into the reaction system when pipetting each reagent.

Assay protocol

▲ Plate set up

	1	2	3	4	5	6	7	8	9	10	11	12
A	A	A	S1	S9	S17	S25	S33	S41	S49	S57	S65	S73
B	B	B	S2	S10	S18	S26	S34	S42	S50	S58	S66	S74
C	C	C	S3	S11	S19	S27	S35	S43	S51	S59	S67	S75
D	D	D	S4	S12	S20	S28	S36	S44	S52	S60	S68	S76
E	E	E	S5	S13	S21	S29	S37	S45	S53	S61	S69	S77
F	F	F	S6	S14	S22	S30	S38	S46	S54	S62	S70	S78
G	G	G	S7	S15	S23	S31	S39	S47	S55	S63	S71	S79
H	H	H	S8	S16	S24	S32	S40	S48	S56	S64	S72	S80

Note: A-H, standard wells; S1-S80, sample wells.

▲ Operating steps

The preparation of standard curve

Dilute 200 $\mu\text{mol/L}$ Standard with double distilled water to a serial concentration. The recommended dilution gradient is as follows: 0, 5, 10, 20, 40, 60, 80, 100 $\mu\text{mol/L}$.

The measurement of samples

- (1) **Standard tube:** Take 0.1 mL of standard solution with different concentrations into numbered 10 mL glass tubes.
Sample tube: Take 0.1 mL of tested sample into numbered 10 mL glass tubes.

- (2) Add 0.1 mL of reagent 1 into each tube of Step 1.
- (3) Add 4 mL of chromogenic agent into each tube of Step 2.
- (4) Fasten the tube mouth with fresh-keeping film, mix fully, and make a small hole in the film. Then incubate the tubes at 100°C for 60 min.
- (5) Cool the tubes to room temperature with running water, centrifuge the tubes at 1600 g for 10 min.
- (6) Take 0.25 mL of the supernatant to the microplate with a micropipette (the precipitation cannot be added to the microplate).
- (7) Measure the OD value at 532 nm with microplate reader.

▲ Operation table

	Standard	Sample
Standard solution of different concentrations (mL)	0.1	
Sample (mL)		0.1
Reagent 1 (mL)	0.1	0.1
Chromogenic agent (mL)	4	4
Fasten the tube mouth with fresh-keeping film, mix fully, and make a small hole in the film. Then incubate the tubes at 100°C for 60 min. Cool the tubes to room temperature with running water, centrifuge the tubes at 1600 g for 10 min. Take 0.25 mL the supernatant to the Microplate with a micropipette. Determine the OD value at 532 nm with Microplate Reader.		

▲ Calculation

Plot the standard curve by using OD value of standard and correspondent concentration as y-axis and x-axis respectively. Create the standard curve with graph software (or EXCEL). The concentration of the sample can be calculated according to the formula based on the OD value of sample.

The standard curve is: $y = ax + b$.

1. Serum (plasma):

$$\text{TBARS } (\mu\text{mol/L}) = (\Delta A - b) + a \times f$$

2. Tissue sample:

$$\text{TBARS } (\mu\text{mol/gprot}) = (\Delta A - b) + a \times f + C_{pr}$$

Note:

y: The absolute OD value of standard ($OD_{\text{Standard}} - OD_{\text{Blank}}$)

x: The concentration of standard

a: The slope of standard curve

b: The intercept of standard curve

f: Dilution factor of sample before test

C_{pr} : Concentration of protein in sample (gprot/L)

ΔA : Absolute OD ($OD_{\text{Sample}} - OD_{\text{Blank}}$)

▲ Notes

1. This kit is for research use only.
2. Instructions should be followed strictly, changes of operation may result in unreliable results.
3. The validity of kit is 6 months.
4. Do not use components from different batches of kit.

Appendix I Performance characteristics

Appendix I Performance characteristics			
Detection range	2.6-100 $\mu\text{mol/L}$	Average intra-assay CV (%)	4.3
Sensitivity	0.85 $\mu\text{mol/L}$	Average inter-assay CV (%)	7.1
Average recovery rate (%)	100.9		

▲ Example analysis

Take 0.1 mL of 10% mouse tissue homogenate, carry the assay according to the operation table.

The results are as follows:

standard curve: $y = 0.0038x - 0.0013$, the average OD value of the sample well is 0.106, the average OD value of the blank well is 0.047, the concentration of protein in sample is 15.18 gprot/L , and the calculation result is:

$$\begin{aligned} \text{TBARS content } (\mu\text{mol/gprot}) &= (0.106 - 0.047 + 0.0013) \div 0.0038 \div 15.15 \\ &= 1.05 \mu\text{mol/gprot} \end{aligned}$$

Appendix II Sample preparation

The following sample pretreatment methods are for reference only.

▲ Serum

Collect fresh blood and stand at 25°C for 30 min to clot the blood. Then centrifuge at 2000 g for 15 min at 4°C . Take the serum (which is the upper light yellow clarified liquid layer) to preserve it on ice for detection. If not detected on the same day, the serum can be stored at -80°C for a month.

▲ Plasma

Take fresh blood into the tube which has anticoagulant, centrifuge at 700-1000 g for 10 min at 4°C . Take the plasma (which is the upper light yellow clarified liquid layer, don't take white blood cells and platelets in the middle layer) to preserve it on ice for detection. If not detected on the same day, the plasma can be stored at -80°C for a month.

▲ Tissue sample

Take 0.02-1g fresh tissue to wash with homogenization medium at 2-8°C . Absorb the water with filter paper and weigh. Homogenize at the ratio of the volume of homogenized medium (2-8°C) (mL): the weight of the tissue (g) =9:1, then centrifuge the tissue homogenate for 10 min at 1500 g at 4°C . Take the supernatant to preserve it on ice for detection. Meanwhile, determine the protein concentration of supernatant (E-BC-K318-M). If not detected on the same day, the tissue sample (without homogenization) can be stored at -80°C for a month.

Note:

1. Homogenized medium: PBS (0.01 M, pH 7.4).
2. Homogenized method:
 - (1) Hand-operated: Weigh the tissue and mince to small pieces (1 mm^3), then put the tissues pieces to glass homogenized tube. Add homogenized medium into homogenized tube, place the tube into the ice bath with left hand, and insert the glass tamping rod vertically into the homogenized tube with the right hand to grind up and down for 6-8 min.

Or put the tissue into the mortar, and add liquid nitrogen to grind fully. Then add the homogenized medium to homogenize.
 - (2) Mechanical homogenate: Weigh the tissue to EP tube, add the homogenized medium to homogenize the tissue with homogenizer instrument (60 Hz, 90s) in the ice bath. (For samples of skin, muscle and plant tissue, the time of homogenization can be properly prolonged.)
 - (3) Ultrasonication: Treat the cells with ultrasonic cell disruptor (200 W, 2 s/ time, interval for 3 s, the total time is 5 min).

▲ Notes for sample

1. Please predict the concentration before assaying. If the sample concentration is not within the range of the standard curve, users must determine the optimal sample dilutions for their particular experiments.
2. If the sample type is not included in the manual, a preliminary experiment is suggested to verify the validity.
3. If a lysis buffer is used to prepare tissue or cell homogenates, there is a possibility of causing a deviation due to the introduced chemical substance.