

Uric Acid (UA) Colorimetric Assay Kit

Catalog No: E-BC-K016-M

Method: Colorimetric method

Specification: 96T (Can detect 80 samples without duplication)

Measuring instrument: Microplate reader

Sensitivity: 1.30 mg/L

Detection range: 1.30-80 mg/L

- ▶ Please kindly provide us the lot number (on the outside of the box) of the kit for more efficient service.

General information

▲ Intended use

This kit can be used to measure the uric acid (UA) content in serum, plasma, urine samples.

▲ Background

Uric acid, a purine metabolite, is degraded into allantoin by uric acid enzymes in most mammals. Due to the absence of uric acid oxidase gene, uric acid is the final product of purine metabolism in humans, so the level of uric acid in human blood is higher than that in most mammals. Uric acid is a physiologically important plasma antioxidant that effectively protects biological targets from the oxidation of hydroxyl radicals, hypochloric acid and peroxynitrite.

▲ Detection principle

Uric Acid in protein-free filtrate reduce phosphotungstic acid to form tungsten blue, allantoin and carbon dioxide, the depth of blue color is proportional to the concentration of uric acid. Uric acid content can be calculated by measuring the OD value at 690 nm.

▲ Kit components & storage

Item	Component	Specification	Storage
Reagent 1	1 g/L Uric Acid Standard	1 mL × 1 vial	2-8°C , 3 months
Reagent 2	Protein Precipitator	30 mL × 1 vial	2-8°C , 3 months
Reagent 3	Alkali Reagent	6 mL × 1 vial	2-8°C , 3 months
Reagent 4	Phosphotungstic Acid Reagent	6 mL × 1 vial	2-8°C , 3 months, shading light
	Microplate	96 wells	No requirement
	Plate Sealer	2 pieces	
<p>Note: The reagents must be stored strictly according to the preservation conditions in the above table. The reagents in different kits cannot be mixed with each other.</p>			

▲ Materials prepared by users



Instruments

Microplate reader (680-700 nm), Micropipettor, Centrifuge, Vortex mixer



Consumptive material

Tips (10 μ L, 200 μ L, 1000 μ L), EP tubes (1.5 mL)



Reagents

Double distilled water, Normal saline (0.9% NaCl), PBS (0.01 M, pH 7.4)

▲ Safety data

Some of the reagents in the kit contain dangerous substances. It should be avoided to touch the skin and clothing. Wash immediately with plenty of water if touching it carelessly. All the samples and waste material should be treated according to the relevant rules of laboratory's biosafety.

▲ Precautions

Before the experiment, please read the instructions carefully, and wear gloves and work clothes.

▲ The key points of the assay

1. The supernatant after centrifugation must be clarified.
2. The color stability of uric acid is poor, so it is recommended to complete colorimetric analysis within 20 min after color development.

Pre-assay preparation

▲ Sample preparation

The samples should be prepared as conventional methods. Also please refer to appendix II.

▲ Dilution of sample

It is recommended to take 2~3 samples with expected large difference to do pre-experiment before formal experiment and dilute the sample according to the result of the pre-experiment and the detection range (1.30-80 mg/L).

The recommended dilution factor for different samples is as follows (for reference only):

Sample type	Dilution factor
Mouse serum	1-2
Rat serum	1
Human serum	1
Porcine serum	1
Dog serum	1-2
Human urine	8-10

Note: The diluent is normal saline (0.9% NaCl) or PBS (0.01 M, pH 7.4).

Assay protocol	
Ambient temperature	25-30°C
Optimum detection wavelength	690 nm

Instructions for the use of transferpettor

- (1) Equilibrate the pipette tip in that reagent before pipetting each reagent.
- (2) Don't add the liquid outside the tips into the reaction system when pipetting each reagent.

Assay protocol

▲ Plate set up

	1	2	3	4	5	6	7	8	9	10	11	12
A	A	A	S1	S9	S17	S25	S33	S41	S49	S57	S65	S73
B	B	B	S2	S10	S18	S26	S34	S42	S50	S58	S66	S74
C	C	C	S3	S11	S19	S27	S35	S43	S51	S59	S67	S75
D	D	D	S4	S12	S20	S28	S36	S44	S52	S60	S68	S76
E	E	E	S5	S13	S21	S29	S37	S45	S53	S61	S69	S77
F	F	F	S6	S14	S22	S30	S38	S46	S54	S62	S70	S78
G	G	G	S7	S15	S23	S31	S39	S47	S55	S63	S71	S79
H	H	H	S8	S16	S24	S32	S40	S48	S56	S64	S72	S80

Note: A-H, standard wells; S1-S80, sample wells.

▲ Operating steps

1. The preparation of standard curve

Dilute 1 g/L uric acid standard with double distilled water to a serial concentration. The recommended dilution gradient is as follows: 0, 10, 20, 30, 40, 50, 60, 80 mg/L.

2. The measurement of samples

- 1) **Standard tube:** add 25 μ L of standard with different concentrations into the tubes.
Sample tube: add 25 μ L of sample into the tubes.
- 2) Add 250 μ L of reagent 2 to each tube and mix fully with the vortex mixer.
- 3) Stand the tubes for 5 min. Centrifuge at 2000 g for 5 min (The supernatant should be clarified).

- 4) Take 160 μL of the supernatant to the corresponding wells of microplate.
- 5) Add 50 μL of reagent 3 and 50 μL of reagent 4 orderly. Mix fully with microplate reader for 10 s and stand at room temperature for 15 min.
- 6) Measure the OD value of each well at 690 nm with microplate reader.
(Note: The color stability of uric acid is poor, so it is suggested to finish the absorbance detection within 20 min.)

▲ Operation table

	Standard tube	Sample tube
Standard with different concentrations (μL)	25	
Sample (μL)		25
Reagent 2 (μL)	250	250
Mix fully with the vortex mixer, stand the tubes for 5 min. Centrifuge at 2000 g for 5 min, take 160 μL of the supernatant to the microplate.		
Supernatant (μL)	160	160
Reagent 3 (μL)	50	50
Reagent 4 (μL)	50	50
Mix fully with microplate reader for 10 s and stand at room temperature for 15 min. Measure the OD value of each well at 690 nm with microplate reader.		

▲ Calculation

Plot the standard curve by using OD value of standard and correspondent concentration as y-axis and x-axis respectively. Create the standard curve with graph software (or EXCEL). The concentration of the sample can be calculated according to the formula based on the OD value of sample.

The standard curve is: $y = ax + b$.

$$\text{UA content (mg/L)} = (\Delta_{690} - b) \div a \times f$$

Note:

y: $OD_{\text{standard}} - OD_{\text{Blank}}$.

x: The concentration of standard.

a: The slope of standard curve.

b: The intercept of standard curve.

ΔA_{690} : Absolute OD ($OD_{\text{Sample}} - OD_{\text{Blank}}$)

f: the dilution multiple of tested samples.

▲ Notes

1. This kit is for research use only.
2. Instructions should be followed strictly, changes of operation may result in unreliable results.
3. The validity of kit is 3 months.
4. Do not use components from different batches of kit.

Appendix I Performance characteristics

Appendix I Performance characteristics			
Detection range	1.30-80 mg/L	Average intra-assay CV (%)	2.0
Sensitivity	1.30 mg/L	Average inter-assay CV (%)	4.0
Average recovery rate (%)	96		

▲ Example analysis

Take 25 μL of human serum sample and carry the assay according to the operation table. The results are as follows:

Standard curve: $y = 0.00108x + 0.00366$, the average OD value of the sample well is 0.098, the average OD value of the blank well is 0.037, the calculation result is:

$$\text{UA content(mg/L)} = \frac{(0.098 - 0.037 + 0.00366)}{0.00108} = 59.87\text{mg/L}$$

Appendix II Sample preparation

The following sample pretreatment methods are for reference only.

▲ Serum

Collect fresh blood and stand at 25°C for 30 min to clot the blood. Then centrifuge at 2000 g for 15 min at 4°C . Take the serum (which is the upper light yellow clarified liquid layer) to preserve it on ice for detection. If not detected on the same day, the serum can be stored at -80°C for a month.

▲ Plasma

Take fresh blood into the tube which has anticoagulant (heparin is recommended), centrifuge at 700-1000 g for 10 min at 4°C . Take the plasma (which is the upper light yellow clarified liquid layer, don't take white blood cells and platelets in the middle layer) to preserve it on ice for detection. If not detected on the same day, the plasma can be stored at -80°C for a month.

▲ Urine

Collect fresh urine and centrifuge at 10000 g for 15 min at 4°C Take the supernatant to preserve it on ice for detection. If not detected on the same day, the urine can be stored at -80°C for a month.

▲ Note for sample

1. Please predict the concentration before assaying. If the sample concentration is not within the range of the standard curve, users must determine the optimal sample dilutions for their particular experiments.
2. If the sample type is not included in the manual, a preliminary experiment is suggested to verify the validity.