

7th Edition, revised in April, 2017

(FOR RESEARCH USE ONLY. DO NOT USE IT IN CLINICAL DIAGNOSIS !)

Ractopamine ELISA Kit

Catalog No: E-FS-E012

96T

This manual must be read attentively and completely before using this product.

If you have any problems, please contact our Technical Service Center for help.

Phone: 240-252-7368(USA)240-252-7376(USA)

Email: techsupport@elabscience.com

Website: www.elabscience.com

Please kindly provide us the lot number(on the outside of the box) of the kit for more efficient service.

Test principle

This kit uses Indirect-Competitive-ELISA as the method. It can detect Ractopamine (RAC) in urine, tissues and feed. This kit is composed of Micro ELISA Plate pre-coated with coupled antigen, HRP conjugate, antibody, standard and other supplementary reagents. During the detection, after adding standard or sample solution, RAC in the samples competes with pre-coated coupled antigen on the Micro ELISA Plate for RAC antibody. Then Horseradish Peroxidase (HRP) conjugate is added to each micro plate well, and TMB substrate is added for color development. There is a negative correlation between the OD value of samples and the concentration of RAC. The residual quantity of RAC in the samples can be calculated by comparing the OD of the samples to the standard curve.

Technical indicator

Sensitivity: 0.1ppb(ng/mL)

Reaction mode: 25°C, 30min~ 15min

Detection limit: Urine ---0.1ppb, Tissue (treatment method 1) ---0.4ppb,
Muscle (treatment method 2) ---0.1ppb, Liver (treatment method 2) ---0.2ppb,
Feed---1ppb

Cross-reactivity: Ractopamine---100%; dobutamine---<1%; Clenbuterol ---<0.1%; Albuterol---<0.1%.

Sample recovery rate: Urine ---95% ± 10%, Tissue/feed---90% ± 15%.

Kits components

Item	Specifications
Micro ELISA Plate	96 wells
Standard Solution	1mLeach 0ppb, 0.1ppb, 0.3ppb, 0.9 ppb, 2.7ppb, 8.1ppb.
High Concentrated Standard (Red cap)	100 ppb, 1 mL
HRP Conjugation (Red cap)	5.5mL
Antibody Working Solution (Blue cap)	9mL
Substrate Reagent A (White cap)	6mL
Substrate Reagent B (Black cap)	6mL
Stop Solution (Yellow cap)	6mL
20× Concentrated Wash Buffer (White cap)	40 mL
10× Redissolved Buffer (Yellow cap)	50 mL
Product Description	1 copy

Other supplies required

Instruments: Microplate reader, Printer, Homogenizer, Nitrogen Evaporators, Oscillators, Centrifuge, Graduated pipette, Balance (sensitivity 0.01 g).

High-precision transfer pipette: Single channel (20-200 μL , 100-1000 μL), Multichannel (300 μL).

Reagents: N-hexane, Acetonitrile, Methanol, Anhydrous sodium sulfate.

Experimental preparation

1. Sample pretreatment Notice: Experimental apparatus should be clean and the pipette should be disposable to avoid cross-contamination during the experiment.

2. Solution preparation

Solution 1: Recover-solution

Dilute the $10 \times$ Recover solution with deionized water for re-dissolution of samples. The recover-solution can be stable for 1 month at 4°C .

3. Sample pretreatment procedure

3.1 Pretreatment of urine sample

Take 20 μL clear urine sample for analysis directly (if the urine sample is muddy, it should be filtered or centrifuged at 4000 r/min for 5 min until the urine sample becomes clear). Samples temporarily not used should be frozen.

Note: Sample dilution factor: 1, minimum detection dose: 0.1 ppb.

3.2 Pretreatment of tissue (pretreatment method 1)

Weigh $2 \pm 0.05\text{g}$ of crushed homogenate, add 6 mL recover-solution. Oscillate fully for 2 min, centrifuge at a speed of over 4000 r/min for 10 min (incubate the sample at 85°C for 10 min before centrifugation if there is a high-content of fat in tissue sample). Take 20 μL of the supernatant for analysis.

Note: Sample dilution factor: 4, minimum detection dose: 0.4 ppb.

3.3 Pretreatment of tissue (muscle and liver, pretreatment method 2)

(1) Accurately weigh $2 \pm 0.05\text{g}$ of crushed homogenate, add 8 mL of acetonitrile solution. Oscillate fully for 2 min, centrifuge at a speed of over 4000 r/min for 10 min at room temperature.

(2) Take 5 mL of the supernatant and dry with nitrogen or air at $50-60^\circ\text{C}$

△ Muscle sample:

Add 1 mL recover-solution and oscillate for 30 seconds. Take 20 μL of the supernatant for analysis.

Sample dilution factor: 1, minimum detection dose: 0.1 ppb.

△ Liver sample:

Add 2 mL n-hexane and oscillate, then add 1 mL deionized water. Oscillate and mix for 30 seconds. Centrifuge at a speed of over 4000 r/min for 10 min at room temperature, remove the

upper layer liquid. Take 50 μL of the lower layer liquid and mix with 50 μL of recover-solution. Take 20 μL for analysis.

Note:Sample dilution factor: 2, minimum detection dose: 0.2ppb.

3.4 Pretreatment of feed sample

- (1) Weigh 1 ± 0.05 g of homogenate feed sample, add 10mLmethanol and 5 g Na_2SO_4 .Oscillate for 2 min, centrifuge at a speed of over 4000r/min for 10min at room temperature.
- (2) Take 1 mL of the supernatant and dry with nitrogen or airat 50-60°C.Add 1 mLof recover-solution to dissolve the remaining dry material. Then add 1 mL n-hexane and mix for 30 seconds.Centrifuge for 10 min at a speed of over 4000 r/min at room temperature.
- (3) Take 20 μL of the lower layer liquid for analysis.

Note:Sample dilution factor: 10, minimum detection dose: 1 ppb.

Assay procedure

Centrifuge the sample again after thawing before the assay. Bring all reagents to room temperature before use.**All the reagents should be mixed thoroughly by gently swirling before pipetting. Avoid foaming.**Dilute 40mL of concentrated wash buffer into 800mLwash working buffer with deionized or distilled water.

1. **Number:** number the sample and standard in order (multiplewell), and keep a record of standard wells and sample wells.
2. **Add sample:** add20 μL of Standard, Blank, or Sample per well, then add 50 μL HRP conjugate to each well.Add 80 μL Antibody working solution, cover the plate with sealer we provided, oscillate for 5s gently to mix thoroughly, shading light incubation for 30min at25°C.
3. **Wash:** uncover the sealer carefully, remove the liquid in each well. Immediately add 250 μL of wash working buffer to each well and wash.Repeat wash procedure for 5 times, 30s intervals/time. Invert the plate and pat it against thick clean absorbent paper (If bubbles exist in the wells, clean tips can be used to prick them).
4. **HRP conjugate:**add 100 μL HRP conjugate to each well, shading light incubation for 30min at 25°C.
5. **Wash:** repeat step 3
6. **Color Development:** add 50 μL of substrate solution A to each well, and then add 50 μL of substrate solution B.Gently oscillate for 5s to mix thoroughly. Incubate shading light for 15min at 25°C.
7. **Stop reaction:**add 50 μL of stop solution to each well,gently oscillate and mix fully to stop the reaction.
8. **OD Measurement:** determine the optical density (OD value) of each well at 450 nm with a microplate reader (the 450/630 nm double wavelength is recommended).This step should be finished in 10min after stop reaction.

Result analysis

1. Absorbance% = $A/A_0 \times 100\%$

A: Average absorbance of standard solution or sample

A_0 : Average absorbance of 0 ppb Standard solution

2. Drawing and calculation of standard curve

Create a standard curve by plotting the absorbance percentage of each standard on the y-axis against the log concentration on the x-axis to draw a semi-logarithmic plot. Add the average absorbance value of sample to standard curve to get corresponding concentration. If samples have been diluted, the concentration calculated from the standard curve must be multiplied by the dilution factor.

For this kit, it is more convenient to use professional analysis software for accurate and fast analysis on many samples.

Notes

- (1) Overall OD value will be lower when reagents have not been brought to room temperature before use or room temperature is below 25°C.
- (2) During the washing procedure, if the wells turn dry, it will lead to bad linear standard curve and poor repeatability, move on to the next step immediately after wash.
- (3) Mix thoroughly and wash the plate completely. The consistency of wash procedure can strongly affect the reproducibility of this ELISA kit.
- (4) Micro ELISA plate should be covered by plate sealer. Avoid the reagents to strong light.
- (5) Do not use expired kit and reagents of different batches.
- (6) TMB should be abandoned if it turns color. When OD value of standard (concentration: 0) < 0.5 unit ($A_{450\text{nm}} < 0.5$), it indicates the reagent may be deteriorated.
- (7) Stop solution is caustic, avoid contact with skin and eyes.

Storage and valid period

Storage: Store at 2-8°C. Avoid freeze / thaw cycles.

Valid Period: 1 year, production date is on the packing box.

7th Edition, revised in April, 2017

Copyright ©2017-2018Elabsience Biotechnology Inc.All Rights Reserved
